

ProSling

Patient lifting slings for **pivot** frame lifters



A range of slings designed to lift patients from seated (chair) or reclined (bed or floor) positions.

- Single attachment pivot clips and buckles for added safety and security
- Versatile enough for general patient transfers, toileting or hygiene care
- Manufactured from the highest quality material, conforms to industry standards (AS/NZS ISO 10535:2011)
- Heavy duty, quick dry 3 mm quilted foam padding in leg and head support sections delivering increased patient comfort
- Brightly coloured handles and adjusters, for easy access and visibility
- Nylon resin clip tested to 300 kg
- Suits users to a maximum weight of 205 kg (Small and Medium); and 300 kg (Large)
- Colour coded sling straps provide easy identification of sling size

Will suit the majority of lifter products on the market. Some lift manufacturers make claims that only their brand of sling should be used on their lifter. There is no Australian Government or TGA directive to support these claims.




General Purpose with Head Support

An easy fitting sling with adjustable straps for increased positioning control that suits general lifting from the floor, bed or chair – in either lying or sitting position. In built head support creates a secure transfer for patients with less controlled body movements.



Hygiene

Designed for toilet transfer, this sling suits lifting to and from commode or profiled beds. Increased positioning control greatly assists correct seating. Includes an open area from mid back to mid thigh for easy removal of clothing and hygiene access.



General Purpose with Head Support



DIMENSIONS in mm

SIZE	A WIDTH	B LEG WIDTH	C LEG LENGTH	D DEPTH	E LENGTH	CAPACITY	CODE
Small	900	230	450	790	965	205 kg	SL146191
Medium	1010	230	490	890	1060	205 kg	SL146291
Large	1080	250	480	990	1180	300 kg	SL146391

Hygiene



Small	835	145	455	285	740	205 kg	SL247191
Medium	1000	160	465	300	765	205 kg	SL247291
Large	1170	170	480	310	790	300 kg	SL247391

It is very important to use the correct sized sling and ensure it is properly fitted before attempting to lift. This will ensure the patient feels safe, dignified and comfortable, and allows the carer to be confident of an effective and safe procedure. Please refer to the ProSling Size Selection Chart for guidance on sizing recommendations. This chart is published in the ProSling Range Guide and is available to download from www.novis.com.au

THERAPEUTIC DEVICES SHOULD ONLY BE USED IN ACCORDANCE WITH MANUFACTURERS INSTRUCTIONS AND UNDER THE CONSENT, SUPERVISION AND MANAGEMENT OF A SUITABLY QUALIFIED HEALTH PROFESSIONAL.