

Compatibility Matrix - 2700 6Ch. Cordless Falls Monitor

Suitable for private residences and care homes, its versatility turns a basic room into a highly optioned care and monitoring facility offering multiple solutions in one simple to operate unit, our most powerful Falls Monitor yet.












- Monitor up to 7 Pads or Mats simultaneously
- Works with a single Premium Pad/ Mat, and up to six cordless Pads/ Mats
- Durable ABS plastic enclosure, and soft touch removable protection case and Caregiver Reset Key
- Alarm memory and auto re-arm
- Amber LED visual alarm indicator
- Caregiver secure reset option
- Battery operated or optional external plug pack
- Up to 6 months battery life (nominal)
- Easy to access battery compartment
- Can be connected to a nurse call system
- 3 selectable alarm tones
- Easy thumbwheel adjustable alarm volume
- "Pause" feature allows only caregiver to attend the patient without setting the alarm off
- Alerts for lost nurse call connection, lost AC power supply & low battery










What's in the box:

- 1 x 2700 Cordless Monitor
- 1 x Quick Start Guide
- 1 x 2642 Caregiver Reset Key
- 1 x Durable Soft Case
- 1 x Non-Absorbent Mounting Strap
- 3 x High Quality AA Batteries

	Code	Product	Self Reset	Description
	3515	Cordless Under Mattress Bed Pad	Y	Located UNDER MATTRESS Sends OUT OF BED message when getting out of bed. Resets OUT OF BED when back in bed.
	2629	Cordless Bed Pad	Y	Sends OUT OF Bed message when getting out of bed. Resets OUT OF BED when back in bed.
	2630	Cordless Chair Pad	Y	Sends OUT OF CHAIR message when getting out of chair. Resets OUT OF CHAIR when back on chair.
	2624	Cordless Floor Mat		Initiates alarm when stepped on. Reset on Falls Monitor.

All radio frequency (RF) signal transmission and reception distances quoted are indicative only and based on testing carried out under ideal conditions. Actual transmission and reception distances may vary and be dependent on specific environmental conditions.

	Code	Product		Description
	2719	Cordless Pad Transmitter		Required by Cordless Pads/Mats. Sends a signal to the 2700.
	3516	Premium Under Mattress Bed Pad		Positioned UNDER mattress. The monitor alarms when getting out of bed. The monitor auto resets when back in bed.
	2626	Premium Bed Pad		Positioned ABOVE mattress. The monitor alarms when getting out of bed. The monitor auto resets when back in bed.
	2628	Premium Wheelchair / Chair Pad	Y	Positioned on chair / wheelchair. The monitor alarms when getting out of chair. The monitor auto resets when back in chair.
	2633	Premium Heavy-Duty Non-Slip Floor Mat		Positioned strategically in an area of risk. The monitor alarms when stepped on.
	2623	SafeTime Bed Pad - 1 Year	Y	Positioned ABOVE mattress. The monitor alarms when getting out of bed. The monitor auto resets when back in bed. The monitor alarms when pad expires.
	2622	SafeTime Bed Pad - 90 days	Y	Positioned on a chair / wheelchair. The monitor alarms when getting out of bed. The monitor auto resets when back in bed. The monitor alarms when pad expires.
	2621	SafeTime Wheelchair / Chair Pad - 1 Year	Y	Positioned on a chair / wheelchair. The monitor alarms when getting out of bed. The monitor auto resets when back in bed. The monitor alarms when pad expires.
	2620	SafeTime Wheelchair / Chair Pad - 90 days	Y	Positioned on a chair / wheelchair. The monitor alarms when getting out of bed. The monitor auto resets when back in bed. The monitor alarms when pad expires.

	Code	Product	Description
	2721	iBeam	Sends alarm message when triggered. Features a narrow detection zone to reduce false alarms, triggers when a patient breaks the beam.
	2588	Personal Emergency Transmitter (PET)	Initiates alarm when RED button pressed. Reset on Falls Monitor.
	3502	ActiveCare Wireless Handset (Includes wall bracket)	Initiates alarm when RED button pressed. Reset on Falls Monitor.
	3503*	Simple Push Button Call Point*	Initiates alarm when pressed. Reset on Falls Monitor.
	3503	ActiveCare Magnetic Reed Switch	Initiates alarm when door is opened. Reset on Falls Monitor.
	2558	Wireless Waterproof Call Button w/pullcord	Initiates alarm when RED button is pressed. Reset on Falls Monitor
	2406	cura1 6V Plug Pack	Optional Plug Pack.
	2642	Caregiver Reset Key	Cancels alarms on Falls Monitor.
	3504	ActiveCare Bracket	Monitor slides into bracket.

All radio frequency (RF) signal transmission and reception distances quoted are indicative only and based on testing carried out under ideal conditions. Actual transmission and reception distances may vary and be dependent on specific environmental conditions.